

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**

Product  
Produit

Power supply for building-in

Name and address of the applicant  
Nom et adresse du demandeur

XP POWER L L C  
Suite 150  
1241 E DYER RD  
Santa Ana, CA 92705 USA

Name and address of the manufacturer  
Nom et adresse du fabricant

XP POWER L L C  
Suite 150  
1241 E DYER RD  
Santa Ana, CA 92705 USA

Name and address of the factory  
Nom et adresse de l'usine

XP POWER (KUNSHAN) LTD  
230 BIN JIANG NAN RD  
ZHANGPU TOWN  
KUNSHAN, 215300 JIANGSU China

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Additional Information on page 2

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Input: 100-240 Vac, 50/60Hz, 2.4 A Max.  
Output: See test report for details.

Trademark (if any)  
Marque de fabrique (si elle existe)



SMT

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

CCB200PSXXYY  
See Page 2

Model / Type Ref.  
Ref. De type

Additional Information on page 2

Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>ème</sup> page

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms  
part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

E146893-20140605 issued on 2014-06-05

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2014-06-12

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

**US-23462-UL**

**Model Details:**

CCB200PSXXYY (where the "XX" can be any number between 12 to 56 indicating main output voltage, "YY" can be SF or blank indicating Single Fuse), may also be provided with additional suffixes "-S", "-C", "-L", and/or "A".

**Factories:**

XP POWER L L C  
990 BENECIA AVE  
US  
SUNNYVALE, CA 94085 United States

**Additional Information:**

Additionally evaluated to EN 60601-1:2006/A1:2013  
National Difference specified in the CB Test Report

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



- UL (US), 333 Pflugsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2014-06-12

Signature:

Jolanta M. Wroblewska



Test Report issued under the responsibility of:



**IEC 60601-1**  
**Medical electrical equipment**  
**Part 1: General requirements for basic safety and essential performance**

Report Reference No.....: E146893-20140605  
 Date of issue .....: 2014-06-05  
 Total number of pages.....: 256

CB Testing Laboratory.....: UL Camas  
 Address .....: 2600 NW Lake Road  
 Camas, WA 98607 USA

Applicant's name.....: XP Power LLC  
 Address .....: 1241 E. Dyer Road, Suite 150  
 Santa Ana, CA 92705 USA

**Test specification:**  
 Standard .....: IEC 60601-1: 2005 + CORR. 1:2006 + CORR. 2:2007 + AM1:2012  
 (or IEC 60601-1: 2012 reprint)  
 Test procedure.....: CB Scheme  
 Non-standard test method.....:

Test Report Form No.....: IEC60601\_1I  
 Test Report Form Originator .....: UL  
 Master TRF .....: 2014-03


**Copyright © 2014 Worldwide System for Conformity Testing and Certification of Electrotechnical Equipment and Components (IECEE), Geneva, Switzerland. All rights reserved.**




This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

If this Test Report Form is used by non-IECEE members, the IECEE/IEC logo and the reference to the CB Scheme procedure shall be removed.

**This report is not valid as a CB Test Report unless signed by an approved CB Testing Laboratory and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.**

**General disclaimer:**  
 The test results presented in this report relate only to the object tested.  
 This report shall not be reproduced, except in full, without the written approval of the Issuing CB testing laboratory. The authenticity of this Test Report and its contents can be verified by contacting the NCB, responsible for this Test Report.

<b>Test item description .....</b>	Power supply for building-in	
<b>Trade Mark .....</b>		
<b>Manufacturer .....</b>	XP Power LLC 1241 E Dyer Road, Suite 150 Santa Ana, CA 92705 USA	
<b>Model/Type reference.....</b>	CCB200PSXXYY, (where the "XX" can be any number between 12 to 56 indicating main output voltage, "YY" can be SF or blank indicating Single Fuse), may also be provided with additional suffixes "-S", "-C", "-L", and/or "A".	
<b>Ratings.....</b>	Input: 100-240 Vac, 50/60Hz, 2.4 A Max. Output: See Model Differences for details.	
<b>Testing procedure and testing location:</b>		
<input type="checkbox"/>	<b>CB Testing Laboratory:</b>	
<b>Testing location/ address .....</b>		
<input type="checkbox"/>	<b>Associated CB Testing Laboratory:</b>	
<b>Testing location/ address .....</b>		
<b>Tested by (name + signature).....</b>		
<b>Approved by (name + signature) .....</b>		
<b>Testing procedure: TMP/CTF Stage 1:</b>		
<b>Testing location/ address .....</b>		
<b>Tested by (name + signature).....</b>		
<b>Approved by (name + signature) .....</b>		
<b>Testing procedure: WMT/CTF Stage 2:</b>		
<b>Testing location/ address .....</b>		
<b>Tested by (name + signature).....</b>		
<b>Witnessed by (name + signature) .....</b>		
<b>Approved by (name + signature) .....</b>		
<input checked="" type="checkbox"/>	<b>Testing procedure: SMT/CTF Stage 3 or 4:</b>	
<b>Testing location/ address .....</b>		
XP Power LLC 1241 E. Dyer Road, Suite 150 Santa Ana, CA 92705 USA		

<b>Tested by (name + signature).....</b>	Rodney Reyes	
<b>Witnessed by (name + signature) .....</b>		
<b>Approved by (name + signature) .....</b>	Tac Pham	
<b>Supervised by (name + signature).....</b>	Bernadette Matsuoka	

**List of Attachments (including a total number of pages in each attachment):**

**National Differences (51 pages)**

**Enclosures (139 pages)**

**Summary of testing:** All the following tests were conducted by XP Power located at Suite 150, 1241 E. Dyer Road, Santa Ana, CA 92705 USA under their SMT status in order to add the optional 5V standby output. All other tests were covered by the IEC 60601-1, 3<sup>rd</sup> Ed investigation covered under CBTC No. US-22682-UL.

**Tests performed (name of test and test clause):**

**Testing location:**

**Power Input (4.11)**

**Humidity Conditioning (5.7)**

**Voltage or Charge Limitation (8.4.3)**

**Leakage Current Tests (8.7)**

**Dielectric Voltage Withstand Test (8.8.3)**

**Temperature (11)**

**Transformer Short Circuit (15.5.1.2)**

**Transformer Overload (15.5.1.3)**

**Summary of compliance with National Differences**

AT, BE, CA, CH, CZ, DE, DK, FI, FR, GB, HU, IL, IT, NL, NO, PL, SE, SG, SI, SK, TR, UA, US

The product fulfils the requirements of IEC 60601-1:2012 (with AM1)

Copy of marking plate

The artwork below may be only a draft. The use of certification marks on a product must be authorized by the respective NCBS that own these marks.



<b>GENERAL INFORMATION</b>	
<b>Test item particulars (see also Clause 6):</b>	
Classification of installation and use .....	For building-in
Device type (component/sub-assembly/ equipment/ system):	Component
Intended use (Including type of patient, application location) :	Component switching power supply
Mode of operation .....	Continuous
Supply connection .....	For building-in
Accessories and detachable parts included.....	None
Other options include .....	None
<b>Testing</b>	
Date of receipt of test item(s) .....	2014-01-15, 2014-03-31
Dates tests performed .....	2014-01-30 to 2014-04-25
<b>Possible test case verdicts:</b>	
- test case does not apply to the test object .....	N/A
- test object does meet the requirement.....	Pass (P)
- test object was not evaluated for the requirement .....	N/E (collateral standards only)
- test object does not meet the requirement.....	Fail (F)
<b>Abbreviations used in the report:</b>	
- normal condition .....	N.C.
- means of Operator protection .....	MOOP
- single fault condition.....	S.F.C.
- means of Patient protection .....	MOPP
<b>General remarks:</b>	
"(See Attachment #)" refers to additional information appended to the report.	
"(See appended table)" refers to a table appended to the report.	
The tests results presented in this report relate only to the object tested.	
This report shall not be reproduced except in full without the written approval of the testing laboratory.	
List of test equipment must be kept on file and available for review.	
Additional test data and/or information provided in the attachments to this report.	
Throughout this report a <input type="checkbox"/> comma / <input checked="" type="checkbox"/> point is used as the decimal separator.	
<b>Manufacturer's Declaration per sub-clause 4.2.5 of IEC60601-1:2012</b>	
The application for obtaining a CB Test Certificate includes more than one factory location and a declaration from the Manufacturer stating that the sample(s) submitted for evaluation is (are) representative of the products from each factory has been provided..... :	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> Not applicable
When differences exist; they shall be identified in the General product information section.	



<b>Name and address of factory (ies)..... :</b>	XP POWER L L C 990 BENECIA AVE US SUNNYVALE CA 94085 UNITED STATES  XP POWER (KUNSHAN) LTD 230 BIN JIANG NAN RD ZHANGPU TOWN KUNSHAN JIANGSU 215300 CHINA
---	---

**General product information:**

The models covered in this report are component power supplies intended for use in Medical Electrical Equipment. They are open frame power supplies intended for building-in.

**Model Differences**

All models in the Model CCB200PSXX-YY Series are identical with exception to the Mains Transformer (T1) and minor secondary components that allow for different output voltage ratings.

**Output Ratings:**

CCB200PS12: 10.1Vdc to 13.5Vdc, 16.7A Max., 200 W Max.

CCB200PS15: 13.6Vdc to 17Vdc, 13.3A Max. 200 W Max.

CCB200PS18: 17.1Vdc to 21Vdc, 11.1A Max. 200 W Max.

CCB200PS24: 21.1Vdc to 26Vdc, 8.3A Max. 200 W Max.

CCB200PS28: 26.1Vdc to 31Vdc, 7.1A Max. 200 W Max.

CCB200PS33: 31.1Vdc to 33Vdc, 6.1A Max. 200 W Max.

CCB200PS36: 33.1Vdc to 42Vdc, 5.6A Max. 200 W Max.

CCB200PS48: 42.1Vdc to 54Vdc, 4.2A Max. 200 W Max.

CCB200PS56: 54.1Vdc to 56Vdc, 3.57 A Max. 200 W Max

Suffix "SF" indicates single fuse provided in the line side of the primary.

Units provided with suffix "-C" provided with cover.

Units provided with suffix "-S" provided with screw terminal.

Units provided with suffix "-L" provided with input leads.

Units provided with suffix "-A" provided with 5V Stand-by output rated 5Vdc, 0.5A.

See Enclosure - Miscellaneous 7-02 for max power output based on model, ambient and cover options.

**Additional Information**

The required clearance values have been assessed for suitability up to 5000 m elevation (1.29 correction factor as per Table 8 of IEC 60601-1).

The models covered under this Report were additionally evaluated to IEC 60601-1: Edition 3.1, 2012-08, EN 60601-1:2006/A1:2013.

The need for the additional testing and evaluation shall be determined in the end product investigation

The nameplate markings provided as an Enclosure - Marking Plate are considered representative of the entire series.

The power supply series covered by this report employ 2 MOPP between Primary and Secondary circuits.

Testing to IEC 60601-1-2 was not conducted by UL and no supporting evidence of compliance has been presented. When submitting this Test Report to other Certification Body, the manufacturer is responsible for providing any additional information that the Body may need in order to issue its Mark, including testing for compliance with IEC 60601-1-2.

**Technical Considerations**

- The product was investigated to the following additional standards:: ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10 + A1:2008) (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance) (includes Deviations for United States), CAN/CSA-C22.2 No. 60601-1 (2014) (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance) (includes National Differences for Canada), IEC 60601-1: Edition 3.1, 2012-

08, EN60601-1:2006/A1:2013

- The product was not investigated to the following standards or clauses:: Electromagnetic Compatibility (IEC 60601-1-2), Clause 14, Programmable Electronic Systems, Biocompatibility (ISO 10993-1)
- The degree of protection against harmful ingress of water is:: Ordinary
- The mode of operation is:: Continuous
- The product is suitable for use in the presence of a flammable anesthetics mixture with air or oxygen or with nitrous oxide:: No
- The means of connection to the mains supply is: for building-in, to be determined in end-product

#### **Engineering Conditions of Acceptability**

When installed in an end-product, consideration must be given to the following:

- These components have been judged on the basis of the required spacings in the ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10 +A1:2012) (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance)CAN/CSA-C22.2 No. 60601-1 (2008) + CSA C22.2 No. 60601-1:2014 (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance), which covers the end-use product for which the component was designed, IEC 60601-1, Edition 3.1, EN 60601-1:2006/A1:2013)
- The product was submitted and evaluated for use at the maximum ambient temperature (T<sub>ma</sub>) permitted by the manufacturer's specification of: 70°C at full rated load and 85°C at 50% of rated load. For models provided with the optional 5V standby, the manufacturer's maximum ambient temperature is 50°C. See Enclosure Miscellaneous 7-02 for additional information regarding power output.
- The end-product Electric Strength Test is to be based upon a maximum working voltage of: Primary-Earthed: 240 Vrms, 340 Vpk, Primary-Secondary: 240 Vrms, 446 Vpk. Models with -A suffix, Primary - Earthed: 240Vrms, 241Vpk; Primary- Secondary: 240Vrms, 468Vpk
- The power supply terminals and/or connectors are: Not investigated for field wiring
- The maximum investigated branch circuit rating is: 20 A
- The investigated Pollution Degree is: 2
- Proper bonding to the end-product main protective earthing termination is: Required
- An investigation of the protective bonding terminals has: Not been conducted
- The following input terminals/connectors must be connected to the end-product supply neutral: Input Connector (CON1) N terminal.
- The following magnetic devices (e.g. transformers or inductor) are provided with an OBJY2 insulation system with the indicated rating greater than Class A (105°C): T1, T2 (Class F, 155°C)
- The following end-product enclosures are required: Electrical, Mechanical, Fire

- Suitable disconnect device is to be provided in the end system
- Temperature, Leakage and Dielectric Strength testing shall be considered in the end system and consideration of non-frequency weighted leakage current (clause 8.7.3e) to also be considered as part of the end product.
- Clearance spacing evaluated for 5000 m altitude. Additional consideration maybe necessary in the end-use product.
- Printed Wiring Board rated 130°C
- Units provided with additional suffix “SF”, provided with only one fuse. The need for additional fusing shall be determined as part of the end product.
- Heatsinks are floating and considered live. They should not be accessible in the end-product
- The device shall be installed in compliance with the enclosure, mounting, spacing, casualty, markings, and segregation requirements of the end-use application
- The power supplies without the suffix –A were evaluated as having 2 MOPP between primary-to-secondary for 240Vrms, 446Vpk, 1 MOPP between primary-to-ground for 240Vac and 354Vpk. In addition Models CCB200PSXX, where XX is 12 to 36 only were evaluated for 2 MOPP between secondary to earth for working voltage of 42Vdc and 1 MOPP for a working voltage of 250Vrms between secondary and earth for BF output considerations.
- Models with -A suffix were evaluated for 1 MOPP between Primary - Earth: 240Vrms, 241Vpk; 2 MOPP between Primary- Secondary: 240Vrms, 468Vpk. In addition Models CCB200PSXX–A, where XX is 12 to 36 only were evaluated for 2 MOPP between secondary to earth for working voltage of 42Vdc and 1 MOPP for a working voltage of 250Vrms between secondary and earth for BF output considerations.
- ME Equipment is component for building-in. Applicability of the following is to be determined in End Product Evaluation: 5.9 - Accessibility, 7 - Identification marking and Documents, 8.4.2 - Accessible Parts Including Applied Parts, 8.6 - Protective Earthing, 8.11.1 - Isolation from Supply Mains, 8.11.3 - Power Supply Cords, 9 - Protection against mechanical hazards, 11.3 - Fire Enclosure, 11.8 - Interruption of power supply, 15.3 - Mechanical Strength, 15.4.1 - Construction of Connectors, 15.4.4 - Indicators
- Overcurrent releases of adequate breaking capacity must be employed in the end product.

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**

Product  
Produit

Switching Power Supply

Name and address of the applicant  
Nom et adresse du demandeur

XP POWER LLC  
SUITE 150  
1241 E DYER RD  
SANTA ANA CA 92705 USA

Name and address of the manufacturer  
Nom et adresse du fabricant

XP POWER LLC  
SUITE 150  
1241 E DYER RD  
SANTA ANA CA 92705 USA

Name and address of the factory  
Nom et adresse de l'usine

XP POWER LLC  
990 BENECIA AVE, SUNNYVALE CA 94085  
USA

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Additional Information on page 2

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Input: 100-240 Vac, 50/60Hz, 2.4 Max.  
Output: See Test Report for details.

Trademark (if any)  
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

SMT

Model / Type Ref.  
Ref. De type

CCB200PSXXYY  
See Page 2

Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>ème</sup> page

Additionally evaluated to EN 60950-1: 2006 / A11: 2009 / A1:  
2010 / A12: 2011; National Differences specified in the CB Test  
Report.

Additional Information on page 2

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60950-1(ed.2), IEC 60950-1(ed.2);am1

As shown in the Test Report Ref. No. which forms  
part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

E139109-A120-CB-1 issued on 2014-05-01

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pflugsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2014-05-01  
Original Issue Date: 2013-03-11

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

**US-21070-A2-UL**

**Model Details:**

CCB200PSXXYY (where the "XX" can be any number between 12 to 56 indicating main output voltage, "YY" can be SF or blank indicating Single Fuse), may also be provided with additional suffixes "-S", "-C", "-L", and/or "A".

**Factories:**

XP POWER (KUNSHAN) LTD  
230 BIN JIANG NAN RD, ZHANGPU TOWN, KUNSHAN, JIANGSU 215300  
CHINA

**Additional Information:**

The original report was modified to include the following changes/additions:  
- Add optional stand-by output including suffix "-A" to model designation.

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2014-05-01  
Original Issue Date: 2013-03-11

Signature:

Jolanta M. Wroblewska



Test Report issued under  
the responsibility of:



**TEST REPORT**  
**IEC 60950-1**  
**Information technology equipment - Safety -**  
**Part 1: General requirements**

**Report Reference No** .....: E139109-A120-CB-1  
**Date of issue** .....: 2013-03-11  
**Total number of pages** .....: 39

**CB Testing Laboratory** .....: UL San Jose  
**Address** .....: 455 E. Trimble Rd., San Jose, CA, 95131-1230, USA

**Applicant's name** .....: XP POWER LLC  
SUITE 150  
**Address** .....: 1241 E DYER RD  
SANTA ANA CA 92705  
UNITED STATES

**Test specification:**

**Standard** .....: IEC 60950-1:2005 (2nd Edition); Am 1:2009  
**Test procedure** .....: CB Scheme  
**Non-standard test method** .....: N/A

**Test Report Form No.** .....: IEC60950\_1C  
**Test Report Form originator** .....: SGS Fimko Ltd  
**Master TRF** .....: 2012-08

**Copyright © 2012 Worldwide System for Conformity Testing and Certification of Electrotechnical Equipment and Components (IECEE), Geneva, Switzerland. All rights reserved.**


This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

If this test Report is used by non-IECEE members, the IECEE/IEC logo and the reference to the CB Scheme procedure shall be removed.




**This report is not valid as a CB Test Report unless signed by an approved CB Testing Laboratory and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.**

**General disclaimer**

The test results presented in this report relate only to the object tested. This report shall not be reproduced, except in full, without the written approval of the Issuing CB Testing Laboratory. The authenticity of this Test Report and its contents can be verified by contacting the NCB, responsible for this Test Report.

<b>Test item description</b> .....	Switching Power Supply
Trade Mark .....	
Manufacturer .....	XP POWER LLC SUITE 150 1241 E DYER RD SANTA ANA CA 92705 UNITED STATES
Model/Type reference .....	CCB200PSXXYY, (where the "XX" can be any number between 12 to 56 indicating main output voltage, "YY" can be SF or blank indicating Single Fuse), may also be provided with additional suffixes "-S", "-C", "-L", and/or "A".
Ratings .....	Input: 100-240 Vac, 50/60Hz, 2.4 Max.  Output: See Model Differences for details.



<b>Testing procedure and testing location:</b>	
<input type="checkbox"/> <b>CB Testing Laboratory</b>	Testing location / address..... :
<input type="checkbox"/> <b>Associated CB Test Laboratory</b>	Testing location / address..... :
	Tested by (name + signature) ..... : _____
	Approved by (name + signature) ... : _____
<input type="checkbox"/> <b>Testing Procedure: TMP/CTF Stage 1</b>	Testing location / address..... :
	Tested by (name + signature) ..... : _____
	Approved by (+ signature) ..... : _____
	Testing location / address..... :
<input type="checkbox"/> <b>Testing Procedure: WMT/CTF Stage 2</b>	Testing location / address..... :
	Tested by (name + signature) ..... : _____
	Witnessed by (+ signature)..... : _____
	Approved by (+ signature) ..... : _____
	Testing location / address..... :
<input checked="" type="checkbox"/> <b>Testing Procedure: SMT/CTF Stage 3 or 4</b>	Testing location / address..... :
	Tested by (name + signature) ..... : Rodney Reyes 
	Approved by (+ signature) ..... : Tac Pham 
	Supervised by (+ signature) ..... : David Drewes 
	Testing location / address..... : XP Power LLC, Suite 150, 1241 E. Dyer Rd., Santa Ana, CA 92705 USA
<input type="checkbox"/> <b>Testing Procedure: RMT</b>	Testing location / address..... :
	Tested by (name + signature) ..... : _____
	Approved by (+ signature) ..... : _____
	Supervised by (+ signature) ..... : _____
	Testing location / address..... :

<b>List of Attachments</b>
National Differences (0 pages)
Enclosures (14 pages)
<b>Summary Of Testing</b>
Unless otherwise indicated, all tests were conducted at XP Power LLC, Suite 150, 1241 E. Dyer Rd., Santa Ana, CA 92705 USA.

Tests performed (name of test and test clause)	Testing location / Comments
<p>Guide Information Page - Maximum Output Voltage, Current, and Volt Ampere Measurement (1.2.2.1)</p> <p>Input: Single-Phase (1.6.2)</p> <p>Energy Hazard Measurements (2.1.1.5, 2.1.2, 1.2.8.10)</p> <p>SELV Reliability Test Including Hazardous Voltage Measurements (2.2.2, 2.2.3, 2.2.4, Part 22 6.1)</p> <p>Humidity (2.9.1, 2.9.2, 5.2.2)</p> <p>Determination of Working Voltage; Working Voltage Measurement (2.10.2)</p> <p>Heating (4.5.1, 1.4.12, 1.4.13)</p> <p>Electric Strength (5.2.2)</p> <p>Power Supply Output Short-Circuit/Overload (5.3.7)</p>	
<p><b>Summary of Compliance with National Differences:</b></p> <p>Countries outside the CB Scheme membership may also accept this report.</p> <p>List of countries addressed: AT, BE, BG, BY, CA, CH, CN, CZ, DE, DK, ES, EU, FI, FR, GB, GR, HU, IE, IL, IT, JP, KR, NL, NO, PL, PT, RO, SE, SG, SI, SK, UA, US</p> <p>The product fulfills the requirements of: CSA C22.2 No. 60950-1-07 + A1:2011, EN 60950-1:2006 + A1:2010 + A11:2009 + A12:2011, UL 60950-1 2nd Ed. Revised 2011-12-19, IEC 60950-1:2005 + A1:2009</p>	

**Copy of Marking Plate** - Refer to Enclosure titled Marking Plate for copy.

**Test item particulars :**

Equipment mobility .....	for building-in
Connection to the mains .....	for building-in
Operating condition .....	continuous
Access location .....	for building-in
Over voltage category (OVC) .....	OVC II
Mains supply tolerance (%) or absolute mains supply values .....	+10%, -10%
Tested for IT power systems .....	Yes
IT testing, phase-phase voltage (V) .....	230
Class of equipment .....	Class I
Considered current rating of protective device as part of the building installation (A) .....	20 A
Pollution degree (PD) .....	PD 2
IP protection class .....	IP X0
Altitude of operation (m) .....	5000
Altitude of test laboratory (m) .....	less than 2000 meters
Mass of equipment (kg) .....	0.394 without cover; 0.582 with cover

**Possible test case verdicts:**

- test case does not apply to the test object ..... : N / A
- test object does meet the requirement ..... : P(Pass)
- test object does not meet the requirement ..... : F(Fail)

**Testing:**

Date(s) of receipt of test item .....	2014-01-15
Date(s) of Performance of tests .....	2014-01-30 to 2014-04-25

**General remarks:**

"(see Enclosure #)" refers to additional information appended to the report.  
 "(see appended table)" refers to a table appended to the report.

Throughout this report a point is used as the decimal separator.

**Manufacturer's Declaration per Sub Clause 4.2.5 of IEC 60950-1:**

Yes

The application for obtaining a CB Test Certificate includes more than one factory and a declaration from the Manufacturer stating that the sample(s) submitted for evaluation is (are) representative of the products from each factory has been provided .....

When differences exist, they shall be identified in the General Product Information section.

**Name and address of Factory(ies):** XP POWER LLC  
 990 BENECIA AVE  
 SUNNYVALE CA 94085  
 UNITED STATES

XP POWER (KUNSHAN) LTD  
230 BIN JIANG NAN RD  
ZHANGPU TOWN  
KUNSHAN  
JIANGSU 215300 CHINA

## GENERAL PRODUCT INFORMATION:

### Report Summary

The original report was modified on 2014-04-30 to include the following changes/additions:  
- Add optional stand-by output including suffix "-A" to model designation.

### Product Description

The models covered in this report are component power supplies intended for use in Information Technology Equipment. They are open frame power supplies intended for building-in.

### Model Differences

All models in the Model CCB200PSXX-YY Series are identical with exception to the Mains Transformer (T1) and minor secondary components that allow for different output voltage ratings. See below for Model Ratings at 70°C Table Below:

#### Output Ratings:

CCB200PS12: 10.1Vdc to 13.5Vdc, 16.7A Max., 200 W Max.  
CCB200PS15: 13.6Vdc to 17Vdc, 13.3A Max. 200 W Max.  
CCB200PS18: 17.1Vdc to 21Vdc, 11.1A Max. 200 W Max.  
CCB200PS24: 21.1Vdc to 26Vdc, 8.3A Max. 200 W Max.  
CCB200PS28: 26.1Vdc to 31Vdc, 7.1A Max. 200 W Max.  
CCB200PS33: 31.1Vdc to 33Vdc, 6.1A Max. 200 W Max.  
CCB200PS36: 33.1Vdc to 42Vdc, 5.6A Max. 200 W Max.  
CCB200PS48: 42.1Vdc to 54Vdc, 4.2A Max. 200 W Max.  
CCB200PS56: 54.1Vdc to 56Vdc, 3.6 A Max. 200 W Max

See Miscellaneous enclosure Power Output Table for additional information regarding power output and the various configurations.

Suffix "SF" indicates single fuse provided in the line side of the primary.

Units provided with suffix "-C" provided with cover.

Units provided with suffix "-S" provided with screw terminal.

Units provided with suffix "-L" provided with input leads.

Units provided with suffix "-A" provided with 5V Stand-by output rated 5Vdc, 0.5A.

### Additional Information

The required clearance values have been assessed for suitability up to 5000 m elevation (1.48 correction factor as per IEC 60664-1, Table A2).

The need for the additional testing and evaluation shall be determined in the end product investigation.

The nameplate markings provided as an Enclosure - Marking Plate are considered representative of the entire series. The word "BETA" on the marking label is not a part of the model designation.

The power supply series covered by this report employ Double/Reinforced Insulation between Primary and Secondary circuits.

### Technical Considerations

- The product was submitted and evaluated for use at the maximum ambient temperature (T<sub>ma</sub>) permitted by the manufacturer's specification of: 70°C at full rated load and 85°C at half rated load. See Miscellaneous enclosure Power Output Table for additional information regarding power output and the various configurations.
- The means of connection to the mains supply is: for building-in, to be determined in end-product., ,
- The product is intended for use on the following power systems: IT, TN
- The equipment disconnect device is considered to be: for building-in, to be determined in end-product.,
- The product was investigated to the following additional standards: EN 60950-1:2006 + A11:2009 + A12:2010 + A12:2011 (which includes all European national differences, including those specified in this test report).

### Engineering Conditions of Acceptability

When installed in an end-product, consideration must be given to the following:

- The following Production-Line tests are conducted for this product: Electric Strength
- The end-product Electric Strength Test is to be based upon a maximum working voltage of: Primary-Earthed: 244 Vrms, 353 Vpk, Primary-SELV: 338 Vrms, 562 Vpk,
- The following secondary output circuits are SELV: All outputs, except model CCB200PS56.
- The following secondary output circuits are at hazardous energy levels: All Outputs
- The power supply terminals and/or connectors are: Not investigated for field wiring
- The maximum investigated branch circuit rating is: 20 A
- The investigated Pollution Degree is: 2
- Proper bonding to the end-product main protective earthing termination is: Required
- An investigation of the protective bonding terminals has: Not been conducted
- The following input terminals/connectors must be connected to the end-product supply neutral: Input Connector (CON1) N terminal.
- The following magnetic devices (e.g. transformers or inductor) are provided with an OBJY2 insulation system with the indicated rating greater than Class A (105°C): T1, T2 (Class F, 155°C),
- The following end-product enclosures are required: Electrical, Mechanical, Fire
- Suitable disconnect device is to be provided in the end system. --
- Temperature, Leakage and Dielectric Strength testing shall be considered in the end system. --
- Clearance spacing evaluated for 5000 m altitude. Additional consideration maybe necessary in the end-use product. --

- Printed Wiring Board rated 130°C. --
- The equipment is provided with a fuse in both the Line and Neutral of the primary circuit. The need for a marking warning service person of the hazards associated with neutral fusing shall be considered in the end-product. --
- Heatsinks are floating and considered live. They should not be accessible in the end-product. --
- Heating test was not conducted on unit with input/output leads. If unit is provided with input and/or output leads, then temperature on leads must be measured and cannot exceed 105°C. --

Abbreviations used in the report:

- normal condition .....	N.C.	- single fault condition .....	S.F.C
- operational insulation .....	OP	- basic insulation .....	BI
- basic insulation between parts of opposite polarity:	BOP	- supplementary insulation .....	SI
- double insulation .....	DI	- reinforced insulation .....	RI

Indicate used abbreviations (if any)