



Ref. Certif. No.

**US-25627-UL**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE**

**CERTIFICAT D'ESSAI OC**

Product  
Produit

Component DC-to-DC Converter

Name and address of the applicant  
Nom et adresse du demandeur

XP POWER L L C  
15641 RED HILL AVE, SUITE 100  
TUSTIN CA 92780 USA

Name and address of the manufacturer  
Nom et adresse du fabricant

XP POWER L L C  
15641 RED HILL AVE, SUITE 100  
TUSTIN CA 92780 USA

Name and address of the factory  
Nom et adresse de l'usine

9 Keji 2nd Rd Technology Industrial Park Tainan 70955  
TAIWAN

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>eme</sup> page

Additional Information on page 2

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

4.5 Vdc - 9 Vdc or 9Vdc to 18Vdc

Trademark (if any)  
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur

SMT

Model / Type Ref.  
Ref. De type

IMM02xxDyyy, IMM02xxSyyy  
See Page 2

Additional information (if necessary may also be reported on page 2)  
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2<sup>eme</sup> page

Additionally evaluated to EN 60601-1:2006 / A1:2013; National Differences specified in the CB Test Report

Additional Information on page 2

A sample of the product was tested and found to be in conformity with  
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

E321744-D1004-1-ULCB issued on 2015-07-16

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Ppingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M7R 3B4 Ontario, CANADA

Date: 2015-07-22

Signature:

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Jolanta M. Wroblewska



Ref. Certif. No.

**US-25627-UL**

Model Details:  
IMM02xxDyyy, IMM02xxSyyy (where xx is 05 or 12 representing input voltage; yyy is 03, 3V3, 05, 12 or 15 representing output voltage)

Additional Information:  
Risk Management was not included in this investigation.


**Additional information (if necessary)**  
**Information complémentaire (si nécessaire)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2015-07-22

Signature:   
Jolanta M. Wroblewska