



Ref. Certif. No.

US-43648-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching Power Supply

Name and address of the applicant

XP POWER L L C
340 Commerce, Suite 100 Irvine, CA 92602
United States

Name and address of the manufacturer

XP POWER L L C
340 Commerce, Suite 100 Irvine, CA 92602
United States

Name and address of the factory

XP Power LLC
455 E. Trimble Road San Jose, CA 95131
United States

Note: When more than one factory, please report on page 2

☒ [Additional Information on page 2](#)

Ratings and principal characteristics

Input Rated: 100-240 Vac, 50/60 Hz, 3 A
Output Rated: See CB Test Report for details.

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

AHM250PSXX-ZZ, AHM250PSXXT-ZZ

☒ [Additional Information on page 2](#)

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to: EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021.
National Differences: EU Group Differences, CA, JP, US
☐ [Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

E146893-D6002-CB-1 issued on 2024-04-05

This CB Test Certificate is issued by the National Certification Body



- ☒ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL Solutions (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-04-05

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-43648-UL

Factory(ies):

XP POWER (KUNSHAN) LTD
230 BIN JIANG NAN RD
ZHANGPU TOWN KUNSHAN, Jiangsu Sheng, 215300
China

XP POWER (VIETNAM) CO LTD
LOT D - 4Q - CN
My Phuoc 3 Industrial Park
Ben Cat District Binh Duong, Binh Duong,
Vietnam

XP Power
Horseshoe Park Pangbourne RG8 7JW
United Kingdom

Additional Model Detail(s):

AHM250PSXXT-ZZ, AHM250PSXX-ZZ, (where XX is any number between 12-48 designating output voltage, ZZ can be blank, "A", "6", or "6A")

Additional information (if necessary)



- ☒ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-04-05

Signature:

Jolanta M. Wroblewska