

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

SWITCHING POWER ADAPTOR

Name and address of the applicant

XP Power Limited
401 Commonwealth Drive
Haw Par Technocentre, Lobby B, #02-02, Singapore 149598,
Singapore

Name and address of the manufacturer

XP Power Limited
401 Commonwealth Drive
Haw Par Technocentre, Lobby B, #02-02, Singapore 149598,
Singapore

Name and address of the factory

Xianglong Road, Huangzhou, New Town District, Shilong Town,
Dongguan, 523326 Guangdong, China

Ratings and principal characteristics

Input: 100-240Vac, 50/60Hz or 50-60Hz or 50Hz or 60Hz,
1.5A; Output: refer to test report 50268660 001

Trademark (if any)

XP

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

AKM90PSXXZZZZYYYYYYYY
#XXX#, #ZZZ# and #Y #are variables, see test reportAdditional information (if necessary may
also be reported on page 2)A sample of the product was tested and
found to be in conformity withIEC 60601-1-6:2010+A1
IEC 60601-1:2005+A1
for national deviations see test reportAs shown in the Test Report Ref. No. which
forms part of this Certificate

50268660 001

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com

Date: 26.08.2019

Signature:

Dipl.-Ing. Univ. S. O. Steinke