



Ref. Certif. No.

US-44273-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Component Switching Power Supply

Name and address of the applicant

XP POWER L L C
340 COMMERCE, SUITE 100 IRVINE, CA 92602
UNITED STATES

Name and address of the manufacturer

XP POWER L L C
340 COMMERCE, SUITE 100 IRVINE, CA 92602
UNITED STATES

Name and address of the factory

XP POWER LLC
455 E. TRIMBLE ROAD SAN JOSE, CA 95131
UNITED STATES

Note: When more than one factory, please report on page 2

☒ [Additional Information on page 2](#)

Ratings and principal characteristics

Input Rated: 100-240 V ac, 50/60 Hz, 1.5 A
Output rated: 3.3, 5, 12, 15, 24, -12 or -15 V dc, Max 40 or 60 W,
Dual or Triple outputs.

☒ [Additional Information on page 2](#)

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

10013486, 10013489, ECM60UT31>2449, ECM60UT31 -XD0166,
ECM60UT31 -XE0410

☒ [Additional Information on page 2](#)

Additional information (if necessary may also be reported on page 2)

National Differences: EU Group Differences, CA, JP, US

☒ [Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012,
IEC 60601-1:2005/AMD2:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

E146893-D1068-1/A0/C0-CB issued on 2024-09-24

This CB Test Certificate is issued by the National Certification Body



- ☒ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-09-24

Signature:

Mauricio Avila



Ref. Certif. No.

US-44273-UL

Factory(ies):

XP POWER (KUNSHAN) LTD
230 BIN JIANG NAN RD
ZHANGPU TOWN KUNSHAN, JIANGSU, 215300
CHINA

XP POWER (VIETNAM) CO LTD
LOT D - 4Q - CN
MY PHUOC 3 INDUSTRIAL PARK
BEN CAT DISTRICT BINH DUONG, BINH DUONG,
VIETNAM

XP POWER
HORSESHOE PARK PANGBOURNE RG8 7JW
UNITED KINGDOM

Additional Model Detail(s):

ECM60UDxx, ECM60UTxx, ECC60UDxx, ECC60UTxx, ECM60UDxx (3X5), ECM60UTxx (3X5),
ECM40UDxx, ECM40UTxx, ECC40UDxx, ECC40UTxx,
where xx can be 21-22, 31-37, all models may be followed by W

Additional Ratings:

For Models ECM60UT31 -XD0166 and 10013486:

Output Rated:

Output 1: 5.6Vdc, 8A

Output 2: 12.5Vdc, 3A

Output 3: -12Vdc, 0.5A

For Models ECM60UT31 -XE0410 and 10013489:

Output Rated:

Output 1: 5Vdc, 2.5 A

Output 2: 12.5Vdc, 3A

Output 3: -12Vdc, 1A

For Model ECM60UT31>2449:

Output Rated:

Output 1: 5Vdc, 2.5 A

Output 2: 12Vdc, 3.0A

Output 3: -12Vdc, 1A

Additionally evaluated to:

EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021

Additional information (if necessary)



- ☒ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-09-24

Signature:

Mauricio Avila