

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

Name and address of the applicant
Nom et adresse du demandeur

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

Model / Type Ref.
Ref. De type

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{ème} page

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

Component Switching Power Supply

XP POWER L L C
Suite 150
1241 E DYER RD
Santa Ana, CA 92705 USA

XP POWER L L C
Suite 150
1241 E DYER RD
Santa Ana, CA 92705 USA

XP Power LLC
990 Benecia Ave, Sunnyvale CA 94085
USA

Additional Information on page 2
See Page 2



SMT

ECM60UT31>2449, ECM60UT31 -XD0166, 10013486,
ECM60UT31 -XE0410, 10013489, See Page 2

Additionally evaluated to EN 60601-1:2006/ A1:2013.
National Differences specified in the CB Test Report.

Additional Information on page 2

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

4786488107-20140919 issued on 2014-09-19

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2014-10-09

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-24045-UL

Model Details:

ECC40UDxx,ECC40UTxx,ECC60UDxx,ECC60UTxx,ECM40UDxx,ECM40UTxx,ECM60UDxx (3X5),
ECM60UDxx,ECM60UTxx (3X5), ECM60UTxx where xx can be 21-22, 31-37
all models may be followed by "W"

Factories:

XP Power (Kunshan) Limited
230, Bin Jiang Nan Road, Zhang Pu Town, Kunshan, Jiangsu 215300
China

Ratings:

Input Rated: 100-240 V ac, 50/60 Hz, 1.5 A
Output rated: 3.3, 5, 12, 15, 24, -12 or -15 V dc, Max 40 or 60 W, Dual or Triple outputs.

For Models ECM60UT31 -XD0166 and 10013486:

Output Rated:
Output 1: 5.6Vdc, 8A
Output 2: 12.5Vdc, 3A
Output 3: -12Vdc, 0.5A

For Models ECM60UT31 -XE0410 and 10013489:

Output Rated:
Output 1: 5Vdc, 2.5 A
Output 2: 12.5Vdc, 3A
Output 3: -12Vdc, 1A

Additional information (if necessary)

Information complémentaire (si nécessaire)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2014-10-09

Signature:

Jolanta M. Wroblewska