



Ref. Certif. No.

**US-42984-UL**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE**

Product

Component Power Supply

Name and address of the applicant

XP POWER L L C  
340 Commerce, Suite 100 Irvine, CA 92602  
United States

Name and address of the manufacturer

XP POWER L L C  
340 Commerce, Suite 100 Irvine, CA 92602  
United States

Name and address of the factory

XP POWER L L C  
900 Benecia Avenue Sunnyvale, CA 94085  
United States

*Note: When more than one factory, please report on page 2*

Ratings and principal characteristics

Models ECS100USxx-By:  
Input Rated: 100-240 V~, 1.9A, 50/60 Hz  
Output Rated: See CB Test Report for details.  
[Additional Information on page 2](#)

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

ECS100US12-B-XE0399, ECS100USxx-By

Additional information (if necessary may also be reported on page 2)

**Additionally evaluated to:** EN 60601-1:2006, EN 60601-1:2006/A1:2013,  
EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021  
National Differences: EU Group Differences, CA, US  
[Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012,  
IEC 60601-1:2005/AMD2:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

E146893-D1038-2/A0/C0-CB issued on 2023-11-16

This CB Test Certificate is issued by the National Certification Body



- ☒ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2023-11-28

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

**US-42984-UL**

**Factory(ies):**

XP POWER (KUNSHAN) LTD  
230 BIN JIANG NAN RD  
ZHANGPU TOWN KUNSHAN, Jiangsu Sheng, 215300  
China

**Additional Ratings:**

Model ECS100US12-B-XE0399:  
Input Rated: 100-240 V~, 1.9A, 50-60 Hz  
Output Rated: 12Vdc, 8.34A

**Additional information (if necessary)**



- ☒ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2023-11-28

Signature:

Jolanta M. Wroblewska