



Ref. Certif. No.

US-44174-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Build in power supply
Name and address of the applicant	XP POWER L L C 340 COMMERCE, SUITE 100 IRVINE, CA 92602 UNITED STATES
Name and address of the manufacturer	XP POWER L L C 340 COMMERCE, SUITE 100 IRVINE, CA 92602 UNITED STATES
Name and address of the factory	XP POWER LLC 455 E. TRIMBLE ROAD SAN JOSE, CA 95131 UNITED STATES
Note: When more than one factory, please report on page 2	<input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	Input Rated: 100-240 V~,50/60 Hz, 3A Output Rated: See Enclosure - Miscellaneous Table of Output Ratings for maximum output details in Test Report.
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	ECS130USxx-yy, ECS130US15-XA1013 <input checked="" type="checkbox"/> Additional Information on page 2
Additional information (if necessary may also be reported on page 2)	National Differences: EU Group Differences, CA, JP, US <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	E146893-D1064-1/A0/C0-CB issued on 2024-08-26

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-08-27

Signature: Mauricio Avila



Ref. Certif. No.

US-44174-UL

Factory(ies):

XP POWER (KUNSHAN) LTD
230 BIN JIANG NAN RD
ZHANGPU TOWN KUNSHAN, JIANGSU, 215300
CHINA

XP POWER (VIETNAM) CO LTD
LOT D - 4Q - CN
MY PHUOC 3 INDUSTRIAL PARK
BEN CAT DISTRICT BINH DUONG, BINH DUONG,
VIETNAM

XP POWER
HORSESHOE PARK PANGBOURNE RG8 7JW
UNITED KINGDOM

Additional Model Detail(s):

ECS130USxx-yy,
(where xx can be any number between 12 and 48 designating the output voltage, yy can be blank, C, or SF).
Models with suffix SF designate single fuse.

Additionally evaluated to:

EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021

Additional information (if necessary)



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-08-27

Signature:

Mauricio Avila