

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

Component Power Supply intended for Building-in

Name and address of the applicant

XP Power LLC
15641 Red Hill Ave., Suite 100
Tustin, CA 92780 USA

Name and address of the manufacturer

XP Power LLC
15641 Red Hill Ave., Suite 100
Tustin, CA 92780 USA

Name and address of the factory

Note: When more than one factory, please report on page 2

XP POWER (KUNSHAN) LTD
230 Bin Jiang Nan Rd., Zhangpu Town,
Kunshan Jiangsu 215321
China Additional Information on page 2

Ratings and principal characteristics

Input: 100-240 Vac, 50/60Hz, 2.5 A; Output:
See Model Differences in Test Report for details.

Trademark (if any)



Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

EPL150PSXX
See Page 2

Additional information (if necessary may also be reported on page 2)

The report was revised to include technical modifications.

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E321744-D1020-1/A2/C0-ULCB issued on 2019-07-31

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-08-08

Signature:

Original Issue Date: 2018-12-06

Jolanta M. Wroblewska



Model Details:

EPL150PSXX (where XX can be any number between 12 and 48 designating the output voltage), may also be provided with suffix "-SF" or "-T" or "-YYYYYY"

Factories:

ABES TECHNOLOGY CO. LTD.
No 78-1 Zhangma St. Xiushui Township CHANGHUA HSIEN 504
TAIWAN

XP POWER (VIETNAM) CO LTD
LOT D - 4Q - CN My Phuoc 3 Industrial Park Ben Cat District Binh Duong
VIETNAM

Additional Information:

Additionally evaluated to EN 60601-1:2006/ A1:2013/ A12:2014; National Differences specified in the CB Test Report.

The original report was modified to include the following changes/additions:

1. Added additional CoA per client's request - "This power supply was evaluated with Two MOPP between Primary and Secondary for 504Vpk/259Vrms; One MOPP primary and Earth for 354Vpk/250Vrms; Two MOPP between Secondary to Ground for working voltage of 48Vdc and 1 MOPP for working voltage of 240Vrms between Secondary and Earth for BF output considerations."
 2. Added isolation item E for secondary voltage of 48Vdc
 3. Corrected testing location from "XP Power Ltd." to XP Power LLC"
- See Test Report.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-08-08
Original Issue Date: 2018-12-06

Signature:

Jolanta M. Wroblewska