



Ref. Certif. No.

US-44143-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Component Power Supply intended for Building-in

Name and address of the applicant

XP POWER L L C  
340 Commerce, Suite 100 Irvine, CA 92602  
United States

Name and address of the manufacturer

XP POWER L L C  
340 Commerce, Suite 100 Irvine, CA 92602  
United States

Name and address of the factory

XP POWER (KUNSHAN) LTD  
230 BIN JIANG NAN RD  
ZHANGPU TOWN KUNSHAN, Jiangsu, 215321  
China

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

Input: 100-240 Vac, 50/60Hz, 2.5 A;  
Output: See Model Differences for details

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

EPL150PSXX, EPL150PS12-XZ1752@

Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed  
National Differences: EU Group Differences, CA, JP, US

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012,  
IEC 60601-1:2005/AMD2:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

E321744-D1036-1/A0/C0-CB issued on 2024-08-15

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-08-15

Signature:

Mauricio Avila



Ref. Certif. No.

**US-44143-UL**

**Factory(ies):**

No 78-1 Zhangma St  
Xiushui Township Changhua Hsien 504  
Taiwan

XP POWER (VIETNAM) CO LTD  
LOT D - 4Q - CN  
My Phuoc 3 Industrial Park  
Ben Cat District, Binh Duong,  
Vietnam

**Additional Model Detail(s):**

EPL150PSXX  
(where XX can be any number between 12 and 48 designating the output voltage), may also be provided with suffix "-SF" or "-T" or "-YYYYYY"

EPL150PS12-XZ1752@  
(where @ is any alphanumeric for marketing purpose only)

**Additional information (if necessary)**



- UL Solutions (US), 333 Pflugsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-08-15

Signature:

Mauricio Avila