




Ref. Certif. No.

US-43802-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Component Power Supply
Name and address of the applicant	XP POWER L L C 340 Commerce, Suite 100 Irvine, CA 92602 United States
Name and address of the manufacturer	XP POWER L L C 340 Commerce, Suite 100 Irvine, CA 92602 United States
Name and address of the factory	XP Power LLC 455 E. Trimble Road San Jose, CA 95131 United States <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	GCS150PSxxyy and GCS150PSxxKyy series Input: 100-240 Vac, 50/60 Hz, 1.8A Output: See CB Test Report for details <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	GCS150PSxxyy, GCS150PSxxKyy, GCS180PSxxyy <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Additional information (if necessary may also be reported on page 2)	National Differences: CA, US <input type="checkbox"/> <a href="#">Additional Information on page 2</a>
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	E146893-D1047-2/A0/C0-CB issued on 2024-05-15

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pflugsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-05-23

Signature:

Mauricio Avila



Ref. Certif. No.

**US-43802-UL**

**Factory(ies):**

XP POWER (KUNSHAN) LTD  
230 BIN JIANG NAN RD  
ZHANGPU TOWN KUNSHAN 215300  
China

XP POWER (VIETNAM) CO LTD  
LOT D - 4Q - CN  
My Phuoc 3 Industrial Park  
Ben Cat District Binh Duong  
Vietnam

XP Power  
Horseshoe Park Pangbourne RG8 7JW  
Great Britain

**Additional Model Detail(s):**

GCS150PSxxyy, GCS150PSxxKyy, GCS180PSxxyy,  
(where xx can be any number between 12 and 48 and yy is "-C", "-TF", "-EF" or blank and may be preceded by "-R", all "-" considered optional, may also be provided with additional suffix "SF" or "S")

**Additional Ratings:**

GCS180PSxxyy series  
Input: 100-240 Vac, 50/60 Hz, 2.2A  
Output: See CB Test Report for details

**Additional information (if necessary)**



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-05-23

Signature:

Mauricio Avila