



Ref. Certif. No.

SG PSB-MD-00594M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	Switching power supply unit (AC-DC Power Supply)
Name and address of the applicant	<b>XP Power LLC.</b> 15641 Red Hill Avenue, Suite 100 Tustin CA 92780 USA
Name and address of the manufacturer	XP Power LLC. 15641 Red Hill Avenue, Suite 100, Tustin CA 92780, USA
Name and address of the factory	Dongguan Cincon Electronics Limited No.1 Jing Xiang Rd., DongCheng Foreign Trade Industrial Park, ZhuShan., DongCheng District, 523128 Dongguan, PEOPLE'S REPUBLIC OF CHINA  Cincon Electronics Co., Ltd. 8-1, Fu Kung Rd., Fu Hsing Park., 506 Fu Hsing Hsiang, Chang Hua Hsien., TAIWAN
Ratings and principal characteristics	See page 2 for ratings.
Customer's Testing Facility (CTF) Stage used	CTF STAGE 2
Model/type Ref.	See page 2 for models.
Additional information (if necessary)	Certificate SG PSB-MD-00594 issued 2022-08-30 is replaced by this version due to technical changes. See test report for National Differences. The risk management requirements of the standard were not addressed. See page 2 for other exclusion clauses.
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012
as shown in the Test Report Ref. No. which forms part of this certificate	7191295801-EEC22

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 057396 0839 Rev. 00

Date, 2022-11-02

( Kim Hock Teo )



## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Trademark / Brand (Image)



Models (continued)

MCE03US03; MCE03US05; MCE03US09; MCE03US12; MCE03US15,  
MCE03US24; MCE03US48  
MCE03US03-P; MCE03US05-P; MCE03US09-P;  
MCE03US12-P; MCE03US15-P, MCE03US24-P;  
MCE03US48-P

Input:

100-240V~, 0.1A, 50/60Hz

Output:

MCE03US03, MCE03US03-P: 3.3VDC, 910mA;  
MCE03US05, MCE03US05-P: 5VDC, 600mA;  
MCE03US09, MCE03US09-P: 9VDC, 333mA;  
MCE03US12, MCE03US12-P: 12VDC, 250mA;  
MCE03US15, MCE03US15-P: 15VDC, 200mA;  
MCE03US24, MCE03US24-P: 24VDC, 125mA  
MCE03US48, MCE03US48-P: 48VDC, 63mA

Exclusion Clauses:

- Clause 11.7 Biocompatibility
- Clause 12.2, 15.1: Usability evaluations
- Clause 17 Electromagnetic Compatibility

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 057396 0839 Rev. 00

Date, 2022-11-02



( Kim Hock Teo )

