

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switching Power Supply
Name and address of the applicant	XP Power Limited 401 Commonwealth Drive Haw Par Technocentre, Lobby B, #02-02, Singapore 149598, Singapore
Name and address of the manufacturer	XP Power Limited 401 Commonwealth Drive Haw Par Technocentre, Lobby B, #02-02, Singapore 149598, Singapore
Name and address of the factory	4F. No.2, Lane 235, Bau-Chiau Road Hsin-Tien Dist., New Taipei City, Taiwan 231
Ratings and principal characteristics	AC 100-240V; 47-63Hz; 5-2A; Class I; for output ratings refer to test report page 10
Trademark (if any)	XP
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	SDM300PSX1yy-z; SDM300PDX2yy-z (X1 = 12, 15, 24, 36 or 48; X2 = 0512 or 1224; yy = 0-9, A-Z and blank; z = blank, E or F; for definition refer to test report)
Additional information (if necessary may also be reported on page 2)	Models differ in model name, o/p construction, o/p circuits/ PCB and chassis, construction of the main transformer T1, optionally used appliance inlet and terminal blocks or connectors; for details see test report
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 + A1 for national deviations see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	50094718 001

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 08.11.2017

Signature: Dipl.-Ing. (FH) A. Klinker