

UK Declaration of Conformity

We

XP POWER LIMITED **Company Name:** Company Address: Horseshoe Park Pangbourne

Berkshire RG8 7JW United Kinadom

Declare that this DoC is issued under our sole responsibility and belongs to the following product.

Product Name: KR series

Product Type: Rack mount power supply

Model Number: All model numbers within the KR series

The object of the declaration above is in conformity to the following UK Regulations and Product Specifications:

Electromagnetic Compatibility EN 61000-6-4:2011, class A Regulations 2016, SI 2016 No. 1091:

CISPR 11, class A IEC 61000-4-3:2010 IEC 61000-4-5:2009

IEC 61000-4-11:2004

EN 61000-6-2:2005

IEC 61000-4-2:2008 IEC 61000-4-6:2013 IEC 61000-4-4:2012

The Restriction of the Use of Certain **Hazardous Substances in Electrical** and Electronic Equipment Regulations

2012, No. 3032:

2011/65/EU

The Electrical Equipment (Safety) Regulations 2016, SI 2016 No. 1101: IEC 61010-1: 2010. 3rd edition

Environmental Conditions: Indoor use

Normal environmental conditions per IEC 61010-1 paragraph 1.4.1

Altitude up to 2000 meters Temperature 5°C to 40° C Humidity 80% maximum Input Mains Fluctuations +/-10%

Pollution Degree 2 per IEC 61010-1, paragraph 3.6.8

Electromagnetic Combability Regulations 2016, SI 2016 No 1091, **Means of Conformity:**

based on the use of testing and Technical Documentation in accordance

with this Regulation.

The Electrical Equipment (Safety) Regulations 2016, SI 2016 No 1101, based on the use of design analysis and testing in accordance with this

Regulation.

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012, No. 3032, based on the use of Technical Documentation in accordance with this Regulation.

1 of 2 102002-247KR Rev. NR, 30 April 2021





Technical Documentation: Technical Document, EMC: 04302021KR.TCF

Technical Document, RoHS II: 04302021KR.TD

Test Labs: NTS

36 Gilbert Street South Tinton Falls, NJ 07701 (USA)

XP POWER LLC 124 West Main Street

High Bridge, NJ 08829-0317 (USA)

EMC Compliance Report: 41733-10.GHV, EV1.5F2.0 EMC (Rev 2)

EMC Compliance Report: 42118-11. EMI-GHV, EX030N075DA (Rev 5)

EMC Compliance Report: EMC/ESD, KR100P030-22 (Rev NR)

EMC Compliance Report: PR062327 Glassman KR 100kV EMC (Rev 2)

Electrical Equipment (Safety) Report: LVD, PS/KR100P030-22

Signed for and on behalf of: XP POWER LLC

PO Box 317

124 West Main Street

High Bridge, NJ 08829-0317 (USA)

Name: Steven DeClario

Signature: Steve DeClario

Function: Director of Engineering

Date: April 30, 2021