

Zertifikat

Certificate



Zertifikat Nr. *Certificate No.*
R 50393359

Blatt *Sheet*
0001

Ihr Zeichen *Client Reference*
164105112

Unser Zeichen *Our Reference*
02-DF-50090556 002

Ausstellungsdatum *Date of Issue*
30.11.2017 (day/mo/yr)

Genehmigungsinhaber *License Holder*
XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Fertigungsstätte *Manufacturing Plant*
030-0000342976

Prüfzeichen *Test Mark*



Geprüft nach *Tested acc. to*
IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1

Zertifiziertes Produkt (Geräteidentifikation)
Certified Product (Product Identification)

Lizenzentgelte - Einheit
License Fee - Unit

Power Supply Switching Power Adaptor

Type Designation : AKM36USXXXC2YYYYYYYY
'XXX'=090-360, 3-digit number which divided by 10 indicates
the output voltage in V, minimum step is 0.1V.
'Y'=0-9,A-Z, '-' or blank; Only for marking purposes
Trademark: XP

Protection class: Class II

Input: AC 100-240V; 50/60Hz or 50-60Hz; 1.0A

Output: 1) DC 9.0V; 4.00A 2) DC 9.1-10V; 3.6A
3) DC 10.1-11V; 3.27A 4) DC 11.1-12V; 3.0A
5) DC 12.1-13V; 2.76A 6) DC 13.1-14V; 2.57A
7) DC 14.1-15V; 2.40A 8) DC 15.1-16V; 2.25A
9) DC 16.1V; 2.23A 10) DC 16.2-17V; 2.11A
11) DC 17.1-18V; 2.0A 12) DC 18.1-19V; 1.89A
13) DC 19.1-20V; 1.8A 14) DC 20.1-21V; 1.71A
15) DC 21.1-22V; 1.63A 16) DC 22.1-23V; 1.56A
17) DC 23.1-24V; 1.5A 18) DC 24.1-25V; 1.44A

Continued on page 0002

6

6

ANLAGE (Appendix): 1

Dem Zertifikat liegt unsere Prüf- und Zertifizierungsordnung zugrunde und es bestätigt die Konformität des Produktes mit den oben genannten Standards und Prüfgrundlagen. Zusätzliche Anforderungen in Ländern, in denen das Produkt in Verkehr gebracht werden soll, müssen zusätzlich betrachtet werden. Die Herstellung des zertifizierten Produktes wird überwacht.
This certificate is based on our Testing and Certification Regulation and states the conformity of the product with the standards and testing requirements as indicated above. Any additional requirements in countries where the product is going to be marketed have to be considered additionally. The manufacturing of the certified product is subject to surveillance.

TÜV Rheinland LGA Products GmbH, Tillystraße 2, 90431 Nürnberg

Tel.: +49 221 806-1371 e-mail: cert-validity@de.tuv.com
Fax: +49 221 806-3935 http://www.tuv.com/safety



Dipl.-Ing. Univ. S. O. Steinke

Zertifikat

Certificate



Zertifikat Nr. *Certificate No.*
R 50393359

Blatt *Sheet*
0002

Ihr Zeichen <i>Client Reference</i>	Unser Zeichen <i>Our Reference</i>	Ausstellungsdatum	Date of Issue
164105112	02-DF-50090556 002	30.11.2017	(day/mo/yr)

Genehmigungsinhaber *License Holder*
XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Fertigungsstätte *Manufacturing Plant*
030-0000342976

Prüfzeichen *Test Mark*



Type Approved
Safety
Regular Production
Surveillance

www.tuv.com
ID 1419052789

Geprüft nach *Tested acc. to*

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1

Zertifiziertes Produkt (Geräteidentifikation)
Certified Product (Product Identification)

Lizenzentgelte - Einheit
License Fee - Unit

Power Supply Switching Power Adaptor

Continuation:

Output:

19) DC 25.1-26V; 1.38A	20) DC 26.1-27V; 1.33A
21) DC 27.1-28V; 1.28A	22) DC 28.1-29V; 1.24A
23) DC 29.1-30V; 1.20A	24) DC 30.1-31V; 1.16A
25) DC 31.1-32V; 1.12A	26) DC 32.1-33V; 1.09A
27) DC 33.1-34V; 1.05A	28) DC 34.1-35V; 1.02A
29) DC 35.1-36V; 1.0A	

Dem Zertifikat liegt unsere Prüf- und Zertifizierungsordnung zugrunde und es bestätigt die Konformität des Produktes mit den oben genannten Standards und Prüfgrundlagen. Zusätzliche Anforderungen in Ländern, in denen das Produkt in Verkehr gebracht werden soll, müssen zusätzlich betrachtet werden. Die Herstellung des zertifizierten Produktes wird überwacht.
This certificate is based on our Testing and Certification Regulation and states the conformity of the product with the standards and testing requirements as indicated above. Any additional requirements in countries where the product is going to be marketed have to be considered additionally. The manufacturing of the certified product is subject to surveillance.

TÜV Rheinland LGA Products GmbH, Tillystraße 2, 90431 Nürnberg

Tel.: +49 221 806-1371 e-mail: cert-validity@de.tuv.com
Fax: +49 221 806-3935 http://www.tuv.com/safety

