



Ref. Certif. No.

DE 2-031242

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

SWITCHING POWER ADAPTOR

Name and address of the applicant

XP Power Limited
401 Commonwealth Drive
Haw Par Technocentre, Lobby B, #02-02 149598, Singapore

Name and address of the manufacturer

XP Power Limited
401 Commonwealth Drive
Haw Par Technocentre, Lobby B, #02-02 149598, Singapore

Name and address of the factory

Xianglong Road, Huangzhou, New Town District, Shilong
Town, Dongguan, 523326 Guangdong, P.R. China

Ratings and principal characteristics

Input: 100-240V~, 50/60Hz or 50-60Hz, 1.0A;
Output: refer to test report

Trademark (if any)

XP

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

AMF36USXXXXXXXXXX
'XXX' and 'Y' are variables, see test report for details.

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

IEC 60601-1-6:2010+A1
IEC 60601-1-11:2015
IEC 60601-1:2005+A1
for national differences see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

CN212W65 001

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TÜV Rheinland LGA Products GmbH
Tillystr. 2, 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web : www.tuv.com

Date: 2021-06-01

Signature:

Dipl.-Ing. Univ. S. O. Steinke