



Product Service

# CERTIFICATE

No. B 057396 0483 Rev. 02

**Holder of Certificate:** **XP Power LLC.**  
340 Commerce, Suite 100  
Irvine CA 92602  
USA

**Certification Mark:**



**Product:** **Power supply**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the Testing, Certification, Validation and Verification Regulations of TÜV SÜD Group have to be complied. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** 7191330037-01-TR

**Valid until:** 2024-10-08

**Date,** 2024-07-10

( Yager Bi )

# CERTIFICATE

No. B 057396 0483 Rev. 02

## Model(s):

**ECF40USxx**

(where xx can be 12, 15, 18, 24, 28, 36 or 48  
designating nominal output voltage)

**Brand Name:** XP

## Parameters:

Rated Input Voltage: 100-240 VAC

Rated Input Current: 1.2 A max

Rated input frequency: 50/60 Hz

Elevation for use: 0-5000 m above sea level

Protection Class: Class I or Class II at end use

Temperature, Ambient: 50°C with 100% rated output  
70°C with 50% rated output

## General Product information:

Models covered are open frame power supplies intended for building-in to be used with Medical Electrical Equipment. Units are intended for used with Class I or Class II end-products.

## Approved models and Rated Outputs:

Model Number	OUTPUT RATING		
	Voltage (VDC )	Max Current (A)	Max Power (W)
ECF40US12	10.1-13.5	3.34	40
ECF40US15	13.6-17	2.67	40
ECF40US18	17.1-21	2.23	40
ECF40US24	21.1-26	1.67	40
ECF40US28	26.1-31	1.43	40
ECF40US36	33.1-42	1.11	40
ECF40US48	42.1-54	0.83	40

# CERTIFICATE

No. B 057396 0483 Rev. 02

## Conditions of Acceptability:

When installing the equipment, all requirements of the standards and the manufacturer's specifications must be met.

## The models require:

- Suitable Mechanical, Fire and Electrical enclosure shall be provided in the end use equipment.
- Proper bonding to the end-product main protective earthing termination is required when the power supply is installed in a Class I end product. When installed in a Class II end product, the power supply shall be mounted in a manner that provides sufficient clearance and creepage distance between the hazardous parts and accessible conductive parts.
- This power supply was evaluated with Two MOPP between Primary and Secondary for 354Vpk/240Vrms; One MOPP primary and Earth for 350Vpk/240Vrms; One MOPP between Secondary to Ground for working voltage of 48Vdc.
- Repeat of leakage current testing and consideration of non-frequency weighted leakage test shall be considered in the end product application.
- This power supply has been evaluated as a continuous operation, ordinary equipment and has not been evaluated for use in the presence of a flammable anesthetic mixture with air, oxygen, or nitrous oxide. The output circuits have not been evaluated for direct patient connection (Type B, BF or CF).
- Protective earthing testing shall be conducted in the end product application.
- Scope of Power Supply evaluation defers the following clauses to the be determined as part of the end product: Clause 7.5 (Safety Signs), Clause 7.9 (Accompanying Documents), Clause 9 (ME Hazard), Clause 10 (Radiation), Clause 14 (PEMS), Clause 16 (ME Systems).
- The product was not investigated to the following standards or clauses: Biocompatibility (ISO 10993-1), Clause 14, Programmable Electronic Systems, Electromagnetic Compatibility (IEC 60601-1-2).

**Tested according to:** EN 60601-1:2006/A12:2014