




VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the product designated below.

Product	DC /DC Power Converter
Model	IMB01 Series
Data Applies To	IMB0105S05; IMB0124S15, IMB0112S05, IMB0124S05
Brand Name	 XP Power
Applicant	XP Power Limited 401 COMMONWEALTH DRIVE HAW PAR TECHNOCENTRE LOBBY B #02-02 149598 SINGAPORE
Manufacturer	N/A
Applicable Standard(s)	EN 55011: 2009+A1: 2010, Class A EN 60601-1-2: 2015 EN 61000-3-2: 2014, EN 61000-3-3: 2013 IEC 61000-4-2: 2008, IEC 61000-4-3: 2006+A1: 20017+A2: 2010, IEC 61000-4-4: 2012, IEC 61000-4-5: 2014+A1: 2017, IEC 61000-4-6: 2013, IEC 61000-4-8: 2009
Reference No.	T190329N09-E
Test Laboratory	Compliance Certification Services Inc. (Tainan Lab.) No.8, Jiucengling, Xinhua Dist., Tainan City 712, Taiwan (R.O.C.)

This device has been tested and found to comply with the stated standard(s), which is(are) required by the Council Medical Directive & Electromagnetic Compatibility Directive (93/42/EEC ; 2014/30/EU). The test results are indicated in the test report and are applicable only to the tested sample identified in the report.



*Jeter Wu / Assistant Manager
Tainan Lab.
Date: May 28, 2019*