



Product Service

CERTIFICATE

No. B 057396 0819 Rev. 00

Holder of Certificate: **XP Power LLC.**
15641 Red Hill Avenue, Suite 100
Tustin CA 92780
USA

Certification Mark:



Product: **Switching power supply unit**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 7191275605-TR

Valid until: 2027-07-11

Date, 2022-07-20

(Yager Bi)



Product Service

CERTIFICATE

No. B 057396 0819 Rev. 00

Model(s): MCE40US03; MCE40US05; MCE40US09; MCE40US12;
MCE40US15; MCE40US24; MCE40US48

MCE40US03-P; MCE40US05-P; MCE40US09-P;
MCE40US12-P; MCE40US15-P; MCE40US24-P;
MCE40US48-P

Brand Name: XP Power

Parameters:

Input:

i) 100-240V~, 0.60-0.25A, 50/60Hz
(for models MCE40US03 and MCE40US03-P only)

ii) 100-240V~, 0.80-0.32A, 50/60Hz
(for models other than MCE40US03 and MCE40US03-P only).

Output:

MCE40US03, MCE40US03-P: 3.3VDC, 9.10A;

MCE40US05, MCE40US05-P: 5VDC, 8.00A;

MCE40US09, MCE40US09-P: 9VDC, 4.44A;

MCE40US12, MCE40US12-P: 12VDC, 3.33A;

MCE40US15, MCE40US15-P: 15VDC, 2.66A;

MCE40US24, MCE40US24-P: 24VDC, 1.66A

MCE40US48, MCE40US48-P: 48VDC, 0.82A

Conditions of Acceptability:

- a) The output was not evaluated as patient connected circuits.
- b) Compliance with the requirements for EMC shall be evaluated for the end use product.
- c) These products have been investigated only as a component part for use in equipment where the suitability of the combination is subject to end product investigation.
- d) Clearance/creepage distance and dielectric strength were evaluated and fulfilled the requirements for MOPP
- e) The leakage current test, clearance/creepage distance and dielectric strength shall be checked in end product.
- f) When installing the equipment, all requirements of the mentioned standard must be fulfilled.
- g) See also Engineering Conditions of Acceptability in CB Test Report.

Tested according to: EN 60601-1:2006/A12:2014