

Certificate



Certificate no.

TA 50223140 01

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-SCL-10034099 001

Test report no.: ZTW1-SCL 10034099 001

Client Reference: 1202012001/VTC

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Adapter)

License Fee - Units

Type Designation	: PDM30USX-YY (XP)	5
X stands for	: 05, 09, 12, 15, 18 or 24	1
Y stands for	: a-z, 0-9 or blank	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 0.6-0.4A	
Output Voltage	: see Appendix	
Output Currents	: see Appendix	
max. Power Output	: 120W	
max. Ambient Temperature	: 40°C	
Protection Class	: II	

This certificate is valid for maximum 5 years.


7

Licensed Test mark:


TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue
(day/mo/yr)
27/02/2012



Dipl.-Ing. W. Hsu



Certificate



Certificate no.

TA 50223140 02

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-SCL-10034099 001

Test report no.: ZTW1-SCL 10034099 003

Client Reference: 31.07.2014

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

17/09/2014

Certificate



Certificate no.

TA 50223140 03

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-SCL-10034099 001

Test report no.: ZTW1- 10034099 004

Client Reference: AB-20150720-A12

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

25/11/2015

Certificate



Certificate no.

TA 50223140 04

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-SCL-10034099 001

Test report no.: ZTW1- 10034099 005

Client Reference: 181002401

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1
EN 60601-1-6:2010+A1
IEC 62366:2007+A1
EN 62366:2008+A1

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

21/03/2018



Dipl.-Ing. (FH) A. Klinker

Appendix to TÜV approved Certificate No.: TA 50223140

Certified Product : Switching Power Adapter
Report Number : 10034099 001
Type Designation : PDM30USX-YY (X=05, 09, 12, 15, 18 or 24;
Y= a-z, 0-9 or blank) (XP)

Model Name	Output Voltage [VDC]	Output Current [A]
PDM30US05-YY	5	4
PDM30US09-YY	9	3
PDM30US12-YY	12	2.5
PDM30US15-YY	15	2
PDM30US18-YY	18	1.65
PDM30US24-YY	24	1.25

Date: 27 Feb., 2012

Certification Body


Dipl.-Ing. W. Hsu

