

Certificate



Certificate no.

TA 50227603 01

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-SCL-10035108 001

Test report no.: ZTW1-SCL 10035108 001

Client Reference: 8288400107/KL180412

Tested to: EN 60601-1:2006
IEC 60601-1:2005

Certified Product: (Switching Power Supply)


License Fee - Units

Type Designation	: SDU150PSXYZ (XP)	5
X stands for	: 12, 24	1
Y stands for	: B or Blank	1
Z stands for	: -D or Blank	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 2-1A	
Output	: 1) 12V, 12.5A max. (models SDU150PS12, SDU150PS12B, SDU150PS12-D, SDU150PS12B-D) 2) 24V, 6.25A max. (models SDU150PS24, SDU150PS24B, SDU150PS24-D, SDU150PS24B-D)	
max. Ambient Temperature:	: 50°C	
Protection Class	: I	

This certificate is valid for maximum 5 years.

8


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)
03/05/2012

Certificate



Certificate no.

TA 50227603 02

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-SCL-10035108 001

Test report no.: ZTW1-BHU 10035108 003

Client Reference: 22.07.2014

Tested to: EN 60601-1:2006+A11
IEC 60601-1:2005

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

W. Hsu
Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

30/09/2014

Certificate



Certificate no.

TA 50227603 03

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-SCL-10035108 001

Test report no.: ZTW1- 10035108 004

Client Reference: ZC-20151210-A1

Tested to: EN 60601-1:2006+A11+A1+A12
IEC 60601-1:2005+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

28/06/2016