



Product Service

CERTIFICATE

No. B 057396 0583 Rev. 00

Holder of Certificate: **XP Power LLC.**
15641 Red Hill Avenue, Suite 100
Tustin CA 92780
USA

Certification Mark:

Product: **Adaptors**
(Open Frame Power Supply)

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 081-191139-000

Valid until: 2024-05-23

Date, 2020-01-02

(Watson Yang)



Product Service

CERTIFICATE

No. B 057396 0583 Rev. 00

Model(s): UCH600PSxx
("xx" can be 12, 24, 36 or 48 to denote different output voltage)

Brand Name: XP Power



Parameters:

Rated input voltage:	100-240 Vac
Rated frequency:	50-60 Hz
Rated input current:	7-2.7 A
Protection class:	I
Max. ambient temperature:	50 °C
Degree of Protection:	IPX0
Rated outputs of models:	See below

Models-#	Rated Outputs
UCH600PS12	12 Vdc, 50 A
	12 Vdc, 0.6 A (for Fan)
	5 Vdc, 1.0 A
UCH600PS24	24 Vdc, 25 A
	12 Vdc, 0.6 A (for Fan)
	5 Vdc, 1.0 A
UCH600PS36	36 Vdc, 16.6 A
	12 Vdc, 0.6 A (for Fan)
	5 Vdc, 1.0 A
UCH600PS48	48 Vdc, 12.5 A
	12 Vdc, 0.6 A (for Fan)
	5 Vdc, 1.0 A

Wat J



Product Service

CERTIFICATE

No. B 057396 0583 Rev. 00

License Conditions:

1. The product provides Means of Patient Protection (MOPP).
2. The product has specified to operate at altitude $\leq 5,000$ m
3. The product does not have direct connection to the patient.
4. Compliance with EMC requirements needs to be evaluated for the end use product.
5. Risk Management was included in this investigation.
6. The following clauses were not part of the manufacturers order and therefore excluded from this testing:
 - 11.7: Biocompatibility of ME EQUIPMENT and ME SYSTEMS
 - 12.2: USABILITY
 - 17: Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS

Tested according to: EN 60601-1:2006/A1:2013

Production 082809
Facility(ies):

Watt J