

Certificate of Compliance

Certificate Number:

[E321744]

Report Reference:

E321744-20240813

Issue Date:

2025-03-18

Issued to:

XP POWER L L C

**340 Commerce, Suite 100, Irvine
CA, 92602, US**

This certificate confirms that representative samples of:

**POWER SUPPLIES, MEDICAL AND DENTAL -
COMPONENT**

See addendum page for Models.

Have been evaluated by UL in accordance with the component requirements in the Standard(s) indicated on this Certificate. UL Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for installation in complete equipment submitted for investigation to UL LLC

See addendum page for Standards.

Additional Information:

See UL Product iQ® at <https://iq.ulprospector.com> for additional information.

This Certificate of Compliance indicates that representative samples of the product described in the certification report have met the requirements for UL certification. It does not provide authorization to apply the UL Recognized Component Mark. Only the Authorization Page that references the Follow-Up Services Procedure for ongoing surveillance provides authorization to apply the UL Mark.

Only those products bearing the UL Recognized Component Mark should be considered as being UL Certified and covered under UL's Follow-Up Services.

Look for the UL Recognized Component Mark on the product.



David Piecuch
UL Mark Certification Program Owner

Certificate of Compliance

Certificate Number: E321744
Report Reference: E321744-20240813
Issue Date: 2025-03-18

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Standards for Safety:

ANSI/AAMI ES60601-1 and ANSI/AAMI ES60601-1:2005 & A1:2012 & A2:2021 (consolidated version) - Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance;

CAN/CSA-C22.2 No. 60601-1 and CAN/CSA-C22.2 No. 60601-1:14 (IEC 60601-1:2005+A1:2012+A2:2020, MOD) (consolidated version) - Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance

Switching Power Supply – Models, CCR300PS12-xxxxxxx [a], CCR300PS15-xxxxxxx [a], CCR300PS24-xxxxxxx [a], CCR300PS28-xxxxxxx [a], CCR300PS30-xxxxxxx [a], CCR300PS48-xxxxxxx [a], CCR300PS54-xxxxxxx [a], CCR200PS12-xxxxxxx [a], CCR200PS15-xxxxxxx [a], CCR200PS24-xxxxxxx [a], CCR200PS28-xxxxxxx [a], CCR200PS30-xxxxxxx [a], CCR200PS48-xxxxxxx [a], CCR200PS54-xxxxxxx [a], CCR420PS12-xxxxxxx [a], CCR420PS15-xxxxxxx [a], CCR420PS24-xxxxxxx [a], CCR420PS28-xxxxxxx [a], CCR420PS30-xxxxxxx [a], CCR420PS48-xxxxxxx [a], CCR420PS54-xxxxxxx [a], CCR550PS12-xxxxxxx [a], CCR550PS15-xxxxxxx [a], CCR550PS24-xxxxxxx [a], CCR550PS28-xxxxxxx [a], CCR550PS30-xxxxxxx [a], CCR550PS36-xxxxxxx [a], CCR550PS48-xxxxxxx [a], CCR550PS54-xxxxxxx [a].